

**DO NOT USE OXIVIR TO CLEAN THE  
NEWTON 7.0 IMAGING SYSTEM.**

**USE ISOPROPYL ALCOHOL WIPES  
TO CLEAN THE NEWTON 7.0  
IMAGING SYSTEM BETWEEN  
COHORTS.**

SC #286  
Eff. 5/23

3.5 x 5